K093387

510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirement of SMDA and 21 CFR 807.92.

1.0 submitter's information

NOV 2 7 2009

Name: Andon Health Co., Ltd.
Address: No 31, Changjiang Road, Nankai District, Tianjin, P.R. China
Phone number: 86-22-6052 6161
Fax number: 86-22-6052 6162
Contact: Liu Yi
Date of Application: 10/23/2009

2.0 Device information

Trade name: Fully Automatic Electronic Blood Pressure Monitor
Common name: Noninvasive blood pressure measurement system
Classification name: Noninvasive blood pressure measurement system

3.0 Classification

Production code: DXN- Noninvasive blood pressure measurement system.
Regulation number: 870.1130
Classification: II
Panel: Cardiovascular

4.0 Predict device information

Manufacturer: Andon Health Co., Ltd.
Device: KD-5961 Fully Automatic Electronic Blood Pressure Monitor
KD-5962 Fully Automatic Electronic Blood Pressure Monitor
KD-5902 Fully Automatic Electronic Blood Pressure Monitor
KD-795 Fully Automatic Electronic Blood Pressure Monitor
510(k) number: K083246 K090771 K083317 K070826

5.0 Device description

KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor is for use by medical professionals or at home and is a non-invasive blood pressure measurement system intended to measure the diastolic and systolic blood pressures and pulse rate of an adult individual

by using a non-invasive technique in which an inflatable cuff is wrapped around the upper arm. The cuff circumference is limited to 22cm-48cm.

It is designed and manufactured according to ANSI/AAMI SP10--manual, electronic or automated sphygmanometers.

The operational principle is based on oscillometric and silicon integrate pressure sensor technology, the result will be shown on a LCD with an electronic interface module, the result can also be classified and displayed by the function of blood pressure classification indicator, If any irregular heartbeat is detected, it can be shown on the LCD.

6.0 Intended use

KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor is for use by medical professionals or at home and is a non-invasive blood pressure measurement system intended to measure the diastolic and systolic blood pressures and pulse rate of an adult individual by using a non-invasive technique in which an inflatable cuff is wrapped around the upper arm. The cuff circumference is limited to 22cm-48cm.

The intended use and the indication for use of KD-5961,KD-5962,KD-5902, as described in its labeling are the same as the predict device KD-5961,KD-5962,KD-5902(K083246 K090771 K083317).

7.0 Summary comparing technological characteristics with predicate device

Technological Characteristics	Comparison result
Design principle	Identical
Appearance	Identical
Patients contact Materials	Identical
Performance	Identical
Biocompatibility	Identical
Mechanical safety	Identical
Energy source	Identical
Standards met	Identical
Electrical safety	Identical
EMC	Identical
Function	Identical

8.0 Performance summary

KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor conforms to the following standards:

- IEC 60601-1, Medical Electrical Equipment - Part 1: General Requirements for Safety, 1988; Amendment 1, 1991-11, Amendment 2, 1995.
- IEC 60601-1-2, Medical Electrical Equipment - Part 1-2: General Requirements for Safety - Collateral standard: Electromagnetic Compatibility - Requirements and Tests (Edition 2:2001 with Amendment 1:2004; Edition 2.1 (Edition 2:2001 consolidated with Amendment 1:2004)).
- AAMI SP10:2002, Manual, electronic or automated sphygmomanometers.
- AAMI / ANSI SP10:2002/A1:2003 --, Amendment 1 to ANSI/AAMI SP10:2002 Manual, electronic, or automated sphygmomanometers.
- AAMI / ANSI SP10:2002/A2:2006 --, Amendment 2 to ANSI/AAMI SP10:2002 Manual, electronic, or automated sphygmomanometers.

9.0 Comparison to the predict device and the conclusion

Our device KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor is substantially equivalent to the Fully Automatic Electronic Blood Pressure Monitor KD-5961,KD-5962,KD-5902 whose 510(k) number is K083246, K090771, K083317.

The several devices are identical in the intended use, the design principle, the material, the appearance, the functions, the performance, the energy source and the applicable standards. KD-5961,KD-5962,KD-5902 only adds a new cuff (cuff circumstance: 22cm-36cm) when compared with the predicted device(K083246, K090771, K083317).

The Environmental parameters of KD-5961 and KD-5902 is changed, the operational range for humidity (<90%) are changed from the predict device whose operational range for humidity is <80%. The pulse rate range is changed from 30-180 times/min to 40-180 times/min.

However, the test in this submission provides demonstration that these small differences do not raise any new questions of safety and effectiveness.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-O66-0609
Silver Spring, MD 20993-0002

Andon Health Co., Ltd.
c/o Mr. Liu Yi
President
No. 31, Changjiang Road, Nankai District
Tianjin, P.R. China, 300193

NOV 2 7 2009

Re: K093387
Trade/Device Name: KD-5961, KD-5962, KD-5902 Fully Automatic Electronic Blood Pressure Monitors
Regulation Number: 21 CFR 870.1130
Regulation Name: Non-invasive blood pressure measurement systems
Regulatory Class: Class II (two)
Product Code: DXN
Dated: October 26, 2009
Received: October 30, 2009

Dear Mr. Yi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must

comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

for Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Statement of Indications for Use

510(k) Number : K093387

Device name: KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor

Indications for use:

KD-5961,KD-5962,KD-5902 Fully Automatic Electronic Blood Pressure Monitor is for use by medical professionals or at home and is a non-invasive blood pressure measurement system intended to measure the diastolic and systolic blood pressures and pulse rate of an adult individual by using a non-invasive technique in which an inflatable cuff is wrapped around the upper arm. The cuff circumference is limited to 22cm-48cm.

Prescription use _____ AND/OR Over-The-Counter Use YES
Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-COUNTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K093387